

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 13 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130116 ✓
Name	Michael Carpenter	ID#	6714 ✓
Date	7-9-10 ✓	Agency	Alaska State Troopers
Phone #	524-3052	Instrument Location	McLorath State Trooper Post
Alco S/N	77111 ✓	Target Value	.078 ✓
High Pressure	700	Alco Test Values	
1 st Alco	.072 ✓	2 nd Alco	.072 ✓
Signature	Mike Carpenter		

COB
7/13/10

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 28th day of July, 2010.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130116

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130116

JULY 09, 2010

OPERATOR'S NAME:

SUPERVISOR

OPERATOR'S NUMBER: 6714

SUBJECT'S LAST NAME:

CARPENTER

SUBJECT'S FIRST NAME/MI :

MICHAEL/R

O.L. #: 0000000

DEPT/AGENCY: MC60

CASE/REPORT: 0000000

TEST TYPE: D

ALCO TARGET VALUE: .078

ALCO S/N: 77111

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.65 in		
ALCO TARGET	.077	13:59
BLANK TEST	.000	14:00
INTERNAL STANDARD	VERIFIED	14:00
ALCO TV 29.65 in	.072	14:01
BLANK TEST	.000	14:02
SUBJECT SAMPLE	.000	14:03
BLANK TEST	.000	14:04
ALCO TV 29.65 in	.072	14:05
BLANK TEST	.000	14:05

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130116

JULY 09, 2010

TIME 13:49

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.65 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~■